



OIML Member State
Japan

OIML Certificate N^o
R76/1992-JP1-04.01
Revision 1

OIML CERTIFICATE OF CONFORMITY

Issuing Authority

Name: National Metrology Institute of Japan /National Institute of Advanced Industrial Science and Technology (NMIJ/AIST)
Address: AIST Tsukuba Central 3-9
Tsukuba Ibaraki 305-8563, Japan
Person responsible: Dr. Tamotsu Nomakuchi, President of AIST

Applicant

Name: SHIMADZU CORPORATION
Address: 1, Nishinokyo-kuwabara-cho, Nakagyo-ku, Kyoto 604-8511, Japan

Manufacturer of the certified pattern

Name: SHIMADZU CORPORATION
Address: 1, Nishinokyo-kuwabara-cho, Nakagyo-ku, Kyoto 604-8511, Japan
Name: SHIMADZU PHILIPPINES MANUFACTURING INC.
Address: Phase 3, Lot 15, Block 15, Cavite EPZ Rosario, Cavite, Philippines

Identification of the certified type:

Non-automatic weighing instruments
Electronic Balance AUW-D/AUW/AUX/AUY series
Further characteristics see page 3

This certificate attests the conformity of the above identified type (represented by the sample or samples identified in the associated Evaluation Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

R76-1, Edition 1992, including amendment 1 (1994),
for accuracy class: **I**

This certificate relates only to the metrological and technical characteristics of the type of instrument covered by the relevant OIML Recommendation identified above.

This certificate does not bestow any form of legal international approval.



OIML Member State
Japan

OIML Certificate N°
R76/1992-JP1-04.01
Revision 1

The conformity was established by the results of tests and examinations provided in the associated:

Test report N° R76/1992-JP1-04.02 that includes 47 pages

Evaluation Report N° 21-007(test report N° 09-04/R76:1992 that includes 17 pages)

The Issuing Authority

The OIML Member

NMIJ/AIST



Dr. T. Nomakuchi

Dr. Y. Miki

President of AIST

2010-01-06

2010-01-06



OIML Member State
Japan

OIML Certificate N°
R76/1992-JP1-04.01
Revision 1

Characteristics:

Type	AUW-D series		AUW series	
Accuracy class	I			
Max	120g ... 220g	600ct ... 1100ct	101g ... 320g	505ct ... 1600ct
e	0.001g	0.01ct	0.001g	0.01ct
d	0.00001g ... 0.0001g	0.0001ct ... 0.001ct	0.00001g ... 0.0001g	0.0001ct ... 0.001ct
n	≤ 220000	≤ 110000	≤ 320000	≤ 160000
Tare-balancing range	100% of Max			
Temperature range	10 °C / 30 °C			
Type	AUX series		AUY series	
Accuracy class	I			
Max	120g ... 320g	600ct ... 1600ct	120g ... 220g	600ct ... 1100ct
e	0.001g	0.01ct	0.001g	0.01ct
d	0.0001g	0.001ct	0.0001g	0.001ct
n	≤ 320000	≤ 160000	≤ 220000	≤ 110000
Tare-balancing range	100% of Max			
Temperature range	10 °C / 30 °C			

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated Evaluation Report is not permitted, although either may be reproduced in full.